



City Government of Tagbilaran

J.A. Clarin Street corner E. Calceta Street
Cogon, Tagbilaran City, Bohol

BIDS AND AWARDS COMMITTEE (BAC)
Landline No.: (038) 422 8049

www.tagbilaran.gov.ph



#AsensoPaMore

SUPPLEMENTAL BID BULLETIN NO. 19-01

**Procurements of Goods
(X-Ray Machine with Computed Radiography High Voltage Generator)
ITB No. 1901-02**


Approved Budget of the Contact – P4,500,000.00

This **SUPPLEMENTAL BID BULLETIN NO.19-01** is being issued to clarify, modify and amend items/specifications in the Bidding Documents in response to the request for clarification from prospective bidders and to confirm key issues addressed during the Pre-Bid Conference held last 17th January 2019 for the aforementioned project:

Submission and Opening of Bids: January 31, 2019 at 2:00 PM	
FROM	TO
Bill of Quantities	
<p><u>Computed Radiography System:</u></p> <p>Single Loader</p> <p>* 20 to 40 MP resolution (87.5 μm)</p> <p>* 70 to 80 plates/hr cycle time</p> <p>* Online UPS with at least 30 minutes back up for the whole system</p> <p>comes with:</p> <p>a. Imaging Plates: 2 unit 14"x17"</p> <p>b. Technologist Workstation (Console)</p> <p>* Branded Desktop PC with Built-in Image pilot Processing Software</p> <p>* 21" Touch Screen Monitor</p> <p>c. Radiologist Workstation</p> <p>* 1 unit Desktop PC with Image Viewer and 2MP Dome Medical</p> <p>Grade Monitor with Software</p> <p>* Storage Capacity at least 2000 images</p> <p>* Hard disk of > 80 GB</p> <p>d. Printer: Continuous ink Tank System</p>	<p><u>Computed Radiography System:</u></p> <p>Single Loader</p> <p>* 20 to 40 MP resolution (87.5 μm)</p> <p>* 70 to 80 plates/hr cycle time</p> <p>* Online UPS with at least 30 minutes back up for the whole system</p> <p>comes with:</p> <p>a. Imaging Plates: 2 unit 14"x17"</p> <p>b. Technologist Workstation (Console)</p> <p>* Branded Desktop PC with Built-in Image pilot Processing Software</p> <p>* 21" Touch Screen Monitor</p> <p>c. Radiologist Workstation</p> <p>* 1 unit Desktop PC with Image Viewer and 2MP Medical</p> <p>Grade Monitor with Software</p> <p>* Storage Capacity at least 2000 images</p> <p>* Hard disk of > 80 GB</p> <p>d. Printer: Continuous ink Tank System</p>

Please be guided accordingly.

Very truly yours,


Atty. KATHY BLANCHE BORJA-JAMILA
 BAC Vice-Chairman