



Republic of the Philippines
City of Tagbilaran
City Hall, Tagbilaran City, 6300 Bohol

INVITATION TO BID FOR Procurement of Goods [X-Ray Machine with Computed Radiography High Voltage Generator]

The City Government of Tagbilaran, through the Continuing Appropriation, intends to apply the sum of **Four Million Five Hundred Thousand Pesos only (Php4,500,000.00)** being the Approved Budget for the Contract (ABC) to payments under the contract for ITB No. 1901-02. Bids received in excess of the ABC shall be automatically rejected.

The City Government of Tagbilaran now invites bids **1 unit X-Ray Machine with Computed Radiography High Voltage Generator – Specs: Control Method: High Frequency Interfer (maximum 60Khz), Rating: 20KW/ 125 KV 160mA/ 100KV 200mA 80KV 250mA, Tube Voltage Setting Range: 40-125KV, etc...** for the **CHO, Tagbilaran City**. Delivery of the Goods is required within **45 Calendar Days** after receipt of Notice to Proceed. Bidders must have completed, within 24 months from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section 11, Instruction to Bidders.

Bidding will be conducted through open competitive procedures using a non-discretionary "pass/fail" criterion as specified in the Implementing Rules and Regulations (IRR) of Republic Act (RA) 9184, otherwise known as the "Government Procurement Reform Act".

Bidding is restricted to Filipino citizens/sole proprietorship, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183 and subject to Commonwealth Act 138.

Interested bidders may obtain further information from the City Government of Tagbilaran at the address given below during 8:00 AM - 5:00 PM Monday to Friday only except holidays.

A complete set of Bidding Documents may be purchased by interested Bidders on **January 9 - 31, 2019** from the address below and upon payment of a non-refundable fee for the Bidding Documents in the amount of **P5,000.00**

It may also be downloaded free of charge from the website of the Procuring Entity (www.tagbilaran.gov.ph), provided that Bidders shall pay the non-refundable fee for the Bidding Documents not later than the submission of their bids.

The City Government of Tagbilaran will hold a **Pre-Bid Conference** on **January 17, 2019, 2:00 P.M. at the City Engineering Conference Room, City Engineering Building, Upper Calceta St., Tagbilaran City**, which shall be open to all interested parties.

Bids must be delivered to the address below on or before **January 31, 2019 at 1:30 PM**. All bids must be accompanied by a bid security in any acceptable forms and in the amount stated in ITB Clause 18.

Bid Opening shall be on **January 31, 2019, 2:00 PM at the City Engineering Conference Room, City Engineering Building, Upper Calceta St., Tagbilaran City**. Bids will be opened in the presence of the Bidders' representatives who choose to attend. Late bids shall not be accepted.

The City Government of Tagbilaran reserves the right to accept or reject any bid, to annul the bidding process and to reject all bids at any time prior to contract award, without thereby incurring any liability to the affected suppliers.

For further information, please refer to:

Ms. Rosemarie P. Palma
Head – BAC Secretariat
City Engineering Building
Upper Calceta Street, Tagbilaran City
Tel. No. (038) 4112072
Fax No. (038) 4228049


CATHELYN O. TORREMOCHA
BAC Chairman

Bill of Quantities

| Item No. | Description | Brand | Quantity | Unit | Unit Cost (Pesos) | Total Cost (Pesos) |
|----------|---|-------|----------|------|-------------------|--------------------|
| 1 | X-RAY MACHINE WITH COMPUTER RADIOGRAPHY HIGH VOLTAGE GENERATOR | | 1 | Unit | | |
| | Specification: | | | | | |
| | * Control Method: High Frequency Inverter (maximum 60Khz) | | | | | |
| | * Rating: 20KW/ 125 KV 160mA/ 100KV 200mA 80KV 250mA | | | | | |
| | * Tube Voltage Setting Range: 40-125KV | | | | | |
| | * Tube Current Setting Range: 80-250mA | | | | | |
| | * Mass Setting Range: 0.3-100mAs (0.32-200mAs) | | | | | |
| | : 125kV 36mAs (71mAs) | | | | | |
| | : 100kV 50mAs (100mAs) | | | | | |
| | : 80kV 63mAs (125mAs) | | | | | |
| | : 45kV 100mAs (200mAs) | | | | | |
| | * Time Setting Range: 0.0025-1.6 sec. (0.0025-3.2 sec) | | | | | |
| | * Anatomical Program Exposure: max 432 user programmable parameter | | | | | |
| | * Display: LED Display | | | | | |
| | * Self Diagnostic Function: Error message indication on the LED Display | | | | | |
| | POWER SOURCE: | | | | | |
| | * System Frequency: Single Phase AC | | | | | |
| | * Nominal Supply Voltage: 100/110/120/200/220/230/240V (50/60Hz) | | | | | |
| | * Transformer Capacity: 1.5kVA | | | | | |
| | X-Ray Tube Unit: | | | | | |
| | * Effective Focus size: 1.2mm | | | | | |
| | * Max. anode heat capacity: 100KJ (140KHU) | | | | | |
| | X-Ray Tube Support Unit: | | | | | |
| | * Focus Movement : Area dose Calculating unit | | | | | |
| | : Vertical Travel focus- floor distance 820-1750mm | | | | | |
| | Operation-manual electromagnetic lock | | | | | |
| | : Longitudinal Travel (along the guide rail)-1050mm | | | | | |
| | Operation-manual electromagnetic lock | | | | | |
| | : Transverse Travel (tower operator)-240mm | | | | | |
| | : X-ray tube rotational axis | | | | | |
| | ±90 degree click stop by 90 degree | | | | | |
| | : Communication of exposure conditions w CR/DR optional | | | | | |
| | Radiographic Table: | | | | | |
| | * Size of Table Top: 18868Lx727Wx550H mm | | | | | |
| | * Table Top Material: wood or soft synthetic Leather (selectable) | | | | | |
| | * X-ray Grid: R=10.1, N=40 lines/cm, fo=100cm Al spacer, fixed (removable grid) | | | | | |
| | * Size of Loadable: 14"x17" to 8 1/2"x8 1/2" | | | | | |
| | cassette (cm or inch to size cassette based on IEC standard 60406) | | | | | |
| | * Max. load of Table: 200kg | | | | | |

Bid Data Sheet

| ITB Clause | |
|------------|---|
| 1.1 | <p>The Procuring Entity is the Local Government Unit of Tagbilaran City, Bohol</p> <p>The name of the Contract is <u>Procurement of Goods [X-Ray Machine with Computed Radiography High Voltage Generator]</u>.</p> <p>The identification number of the Contract is: <u>ITB No. 1901-02.</u></p> |
| 1.2 | <p>The lot(s) and reference is: <u>N/A</u>.</p> |
| <u>2</u> | <p>The Funding Source is:</p> <p>Continuing Appropriation, in the amount of Four Million Five Hundred Thousand Pesos only (Php4,500,000.00)</p> <p>The name of the Project is: <u>ITB No. 1901-02 - Procurement of Goods [X-Ray Machine with Computed Radiography High Voltage Generator]</u>.</p> |
| 3.1 | <p>No further instructions.</p> |
| 5.1 | <p>No further instructions.</p> |
| 5.2 | <p>Foreign bidders, except those falling under ITB Clause 5.2(b), may not participate in this Project.</p> |
| 5.4 | <p>The Bidder must have completed, within the period specified in the Invitation to Bid and ITB Clause 12.1(a)(iii), a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.</p> |
| 7 | <p>No further instructions.</p> |
| 8.1 | <p>Subcontracting is not allowed.</p> |
| 8.2 | <p>Not applicable.</p> |
| 9.1 | <p>The Procuring Entity will hold a Pre-bid Conference for this Project on <u>January 17, 2019, 2:00PM at City Engineering Conference Room, CEO Bldg., Upper Calceta St., Tagbilaran City.</u></p> |
| 10.1 | <p>The Procuring Entity's address is:</p> <p>City Government of Tagbilaran c/o Cathelyn O. Torremocha BAC Chairman Bids and Awards Committee Tagbilaran City</p> <p>Tel. No. 411-2222 loc. 135 Fax No. 4228049 bac2013tagbilaran@yahoo.com.ph</p> |
| 12.1 | <p>Eligibility and Technical Documents must consists of:</p> <p>a. PhilGEPS Certificate of Registration (Platinum)</p> <ul style="list-style-type: none"> - DTI Certificate / SEC - Mayor's Permit - Tax Clearance - Audited Financial Statement (latest) - Annual Income Tax Return with BIR Efps Confirmation of Payment (latest) <p>b. Bidder's statement of all its ongoing government or private</p> |

| | |
|-------------|--|
| | <p>contracts including contracts awarded but not yet started</p> <p>c. Bidder's statement of its Single Largest Completed Contract similar to the contract to be bid</p> <p>d. Net Financial Contracting Capacity or Credit Line Commitment</p> <p>e. JVA (if applicable)</p> <p>f. Bid Security</p> <p>g. Production/Delivery Schedule</p> <p>h. Manpower Requirements</p> <p>i. After-sales service/parts (if applicable)</p> <p>j. Technical Specifications</p> <p>k. Omnibus Sworn Statement</p> |
| 12.1(a)(ii) | The bidder's SLCC similar to the contract to be bid should have been completed within two (2) years prior to the deadline for the submission and receipt of bids. |
| 13.1 | No additional requirements. |
| 13.1(b) | No further instructions. |
| 13.1(c) | No additional requirements. |
| 13.2 | The ABC: Four Million Five Hundred Thousand Pesos only (Php4,500,000.00) Any bid with a financial component exceeding this amount shall not be accepted. |
| 15.4(a)(iv) | No incidental services are required. |
| 15.4(b) | Not applicable. |
| 16.1(b) | The Bid prices for Goods supplied from outside of the Philippines shall be quoted in Philippine Pesos. |
| 16.3 | Not applicable. |
| 17.1 | Bids will be valid until May 31, 2019 . |
| 18.1 | <p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p>1. The amount of not less than Php 90,000 (2% of ABC), if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit;</p> <p>2. The amount of Php 225,000 (5% of ABC), if bid security is in Surety Bond.</p> |
| 18.2 | The bid security shall be valid until May 31, 2019 . |
| 20.3 | Each Bidder shall submit one (1) original and two (2) copies of the first and second components of its bid. |
| 21 | <p>Bids must be delivered to BAC Office, City Engineering Building, Upper Calceta St., Tagbilaran City.</p> <p>The deadline for submission of bids is <u>January 31, 2019 at 1:30 PM</u>.</p> |
| 24.1 | <p>The place of bid opening is <u>City Engineering Conference Room, CEO Bldg., Upper Calceta St., Tagbilaran City.</u></p> <p>The date and time of bid opening is <u>January 31, 2019 at 2:00 PM</u>.</p> |
| 24.2 | No further instructions. |

| | |
|---------|---|
| 27.1 | No further instructions. |
| 28.3 | Grouping and Evaluation of Lots – All items to be grouped together to form one complete Lot that will be awarded to one Bidder to form one complete contract. The goods are grouped in a single lot and the lot shall not be divided into sub-lots for the purpose of bidding, evaluation, and contract award. |
| 28.3(b) | Bid modification is not allowed. |
| 28.4 | No further instructions. |
| 29.2 | No additional requirements. |
| 32.4(f) | No additional requirement. |